

LEL CORRECTION FACTORS EN 60079-20-1:2010 – CALIBRATED TO METHANE = 100

Instrument	Microclip/MAX XT2 & 2				M5				Quattro				Ultra			
	MP 75		MP 75C		4P90		4P90C		4p-75		4p-75c		1LEL75		1LEL75C	
	RS	CF	RS	CF	RS	CF	RS	CF	RS	CF	RS	CF	RS	CF	RS	CF
Hydrogen	143	0,7	144	0,7	188	0,8	116	0,9	134	0,7	136	0,7	126	0,8	126	0,8
Methane	100	1,0	100	1,0	100	1,0	100	1,0	100	1,0	100	1,0	100	1,0	10	1,0
Ethene/ Ethylene	97	1,0			82	1,2	79	1,3			89	1,1				
Acetylene	99	1,0	99	1,0	82	1,2	80	1,3	94	1,1	90	1,1				
Propane	56	1,8	56	1,8	62	1,6	61	1,6	61	1,6	59	1,7	56	1,8	56	1,8
n-Butane	60	1,7	58	1,7	53	1,9	51	2,0	58	1,7	56	1,8	70	1,4	71	1,4
n-Pentane	70	1,4	72	1,4	61	1,6	58	1,7	69	1,4	66	1,5	66	1,5	66	1,5
N-Hexane					45	2,2	43	2,3	55	1,8	53	1,9	46	2,2	48	2,1
Cyclo Hexane	61	1,6			57	1,6	55	1,8	60	1,7	59	1,7				
n-Heptane	63	1,6	61	1,6	57	1,8			62	1,6			62	1,6		
n-Octane	34	2,9			31	3,2			33	3,0			43	2,3		
n-Nonane													34	2,9		
Unleaded Petrol	65	1,5			65	1,5			68	1,5						
Methanol	99	1,0			86	1,2			95	1,1			102	1,0		
Ethanol	83	1,2			77	1,3			89	1,1						
Iso Propyl Alcohol	68	1,5			63	1,6			71	1,4						
Ammonia					144	0,7	139	0,7	159	0,6	164	0,6				
Acetone	69	1,4			69	1,4			74	1,4						
Carbon Monoxide					133	0,9	110	0,9	123	0,8	130	0,8				
Ethyle Acetate	64	1,6			65	1,5			72	1,4						
Methyl Ethyl Ketone	65	1,5			59	1,7			57	1,8						
Toluene	54	1,9			44	2,3			44	2,3						

* - T90 for Ammonia is extended

C – Filtered sensor

RS – Relative sensitivity

CF – Correction factor

The results are intended for guidance only. For the most accurate measurements, an instrument should be calibrated by a certified technician.

